

## Artículo original

# Accuracy of Linear Measurements of Dental Models Scanned Through 3D Scanner and Cone-Beam Computed Tomography in Comparison with Plaster Models

*Exactitud de las mediciones lineales de modelos dentales digitalizados a través de escáner 3D y tomografía computarizada de haz cónico en comparación con modelos de yeso*

*Precisão de medições lineares de modelos dentários digitalizados através de scanner 3D e tomografia computadorizada Cone-Beam em comparação com modelos de gesso*

Léuri Antunes da Silva-Dantas<sup>1</sup> ✉, Amanda Lury Yamashita<sup>1</sup> ✉, Eder Alberto Sigua-Rodriguez<sup>1,2</sup> ✉ CvLAC, Mariliani Chicarelli<sup>1</sup> ✉, Lilian Cristina Vessoni-Iwaki<sup>1</sup> ✉, Liogi Iwaki Filho<sup>1</sup> ✉

1. Department of Dentistry, State University of Maringá, Maringá, Paraná, Brazil.

2. Centro de Investigaciones del Colegio Odontológico (CICO), Institución Universitaria Colegios de Colombia, Bogotá, Colombia.

## Fecha correspondencia:

Recibido: enero de 2019.

Aceptado: julio de 2019.

## Forma de citar:

da Silva-Dantas LA, Yamashita AY, Sigua-Rodriguez EA, Chicarelli M, Vessoni-Iwaki LC, Filho LI. Accuracy of Linear Measurements of Dental Models Scanned Through 3D Scanner and Cone-Beam Computed Tomography in Comparison with Plaster Models. Rev. CES Odont 2019; 32(2): 7-16.

## Open access

© Derecho de autor

Licencia creative commons

Ética de publicaciones

Revisión por pares

Gestión por Open Journal System

DOI: [http://dx.doi.org/10.21615/](http://dx.doi.org/10.21615/cesodon.32.2.1)

cesodon.32.2.1

ISSN 0120-971X

e-ISSN 2215-9185

## Abstract

**Introduction and objective:** Virtual surgical planning uses clinical data, image testing, plaster models of dental arches and clinical photos to simulate an orthognathic. There are two ways to perform the scanning of plaster models: scanning for cone-beam computed tomography (CBCT) or 3D scanner. The purpose of this study was to assess the accuracy and the degree of magnification of plaster model images obtained through 3D scanner and CBCT. **Materials and methods:** The control group was the measurement performed on 40 plaster models by Mitutoyo caliper. The same 40 models were scanned through 3D scanner and CBCT in order to compare the degree of distortion. The models were tested on the Dolphin software. Six measurements were performed in upper and lower arches: intermolar distance; intercanine distance; segment A; segment B; mesiodistal and cervico-incisal distance of the right-side central incisor. **Results:** There was no statistically significant difference for upper and lower models. However, CBCT had the degree of distortion of 2.34%, while the 3D scanner presented the degree of distortion of 2.37% comparing the degree of distortion of both methods with the digital caliper. **Conclusions:** It can be concluded that only the distances of segments A and B of the upper model were not compatible in both scanning methods with the measurements of digital caliper. However, considering all of the measurements, 3D scanner and CBCT are trustworthy to perform

linear measurements on digital models and are sufficiently adequate for initial diagnosis and are clinically acceptable in clinical dental practices.

**Keywords:** Cone-beam computed tomography, orthognathic surgery, dental arch.

## Resumen

**Introducción y objetivo:** La planificación quirúrgica virtual para la simulación de cirugías ortognática utiliza datos clínicos, pruebas de imagen, modelos de yeso y fotos clínicas. Hay dos formas de realizar el escaneo de los modelos de yeso: escaneo con la tomografía computarizada de haz cónico (CBHC) o escáner 3D. El propósito de este estudio fue evaluar la precisión y el grado de alteración de las imágenes del modelo de yeso obtenidas a través del escáner 3D y la CBHC. **Materiales y métodos:** El grupo control fue la medida realizada en 40 modelos de yeso con el calibrador Mitutoyo. Los mismos 40 modelos fueron escaneados a través de un escáner 3D y CBHC para comparar el grado de distorsión. Los modelos fueron evaluados en el software Dolphin. Se realizaron seis medidas en los arcos superior e inferior: distancia intermolar; distancia intercanina; segmento A; segmento B; Distancia mesiodistal y cervico-incisal del incisivo central del lado derecho. **Resultados:** No hubo diferencia estadísticamente significativa para los modelos superiores e inferiores. Sin embargo, la CBHC tuvo un grado de distorsión de 2.34%, mientras que el escáner 3D presentó un grado de distorsión de 2.37% comparando el grado de distorsión de ambos métodos con el calibrador digital. **Conclusión:** Solo las distancias de los segmentos A y B del modelo superior no fueron compatibles en ambos métodos de escaneo con las medidas del calibrador digital. Sin embargo, considerando todas las mediciones, el escáner 3D y la CBCT son confiables para realizar medidas lineales en modelos digitales, son suficientemente adecuados para el diagnóstico inicial y son clínicamente aceptables en las prácticas clínica odontológica.

**Palabras clave:** tomografía computarizada de haz cónico, cirugía ortognática, arco dental.

## Resumo

**Introdução and Objetivo:** O planejamento cirúrgico virtual para a simulação da cirurgia ortognática usa dados clínicos, exames de imagem, modelos de gesso e fotos clínicas. Existem duas maneiras de digitalizar modelos de gesso: digitalização com tomografia computadorizada de feixe cônico (CBFC) ou scanner 3D. O objetivo deste estudo foi avaliar a acurácia e o grau de alteração das imagens do modelo de gesso obtidas através do scanner 3D e do CBFC. **Materiais e métodos:** O grupo controle foi a medida feita em 40 modelos de gesso com o calibrador Mitutoyo. Os mesmos 40 modelos foram digitalizados através de um scanner 3D e CBHC para comparar o grau de distorção. Os modelos foram avaliados no software Dolphin. Seis medições foram realizadas nos arcos superior e inferior: distância intermolar; distância intercanina; segmento A; segmento B; Distância mesiodistal e cervico-incisal do incisivo central do lado direito. **Resultados:** Não houve diferença estatisticamente significativa para os

modelos superior e inferior. No entanto, CBFC tinha um grau de distorção de 2,34%, enquanto o scanner 3D introduziu um grau de distorção de 2,37% na comparação entre o grau de distorção de ambos os métodos com calibrador digital. **Conclusões:** As distâncias dos segmentos A e B do modelo superior não eram compatíveis em ambos os métodos de verificação com medições de calibre digitais. No entanto, considerando-se todas as medições, scanner 3D CBFC e são confiáveis para medições lineares em modelos digitais são adequados o suficiente para o diagnóstico inicial e são clinicamente aceitáveis nas práticas de consultório odontológico.

**Palavras-chave:** tomografia computadorizada de feixe cônico, cirurgia ortognática, arco dentário.

## Introduction

Currently, advances in computer technology provide dental surgeons with the ability to integrate all relevant information of the three-dimensional virtual surgical planning (3D VSP) into a single model of multi-modality imaging, whose clinical advantage is to study complex asymmetric deformities within the orthognathic surgery (1). In a simple way, 3D VSP is the processes that uses clinical data, image testing, and plaster models of dental arches and clinical photos of the patient to simulate an orthognathic surgery in a virtual environment and then transfer the planned movements to the surgical act. In addition, it is used as a communication tool among the orthodontist, the oral and maxillofacial surgeon and the patient, in order to determine the amount and the direction of the surgical movement of hard and soft tissues, as well as assisting in pre- and post-operative orthodontic treatment (2).

The use of 3D images in the maxillofacial area provides accurate diagnostic information based on an accurate visualization of anatomical structures (3,4). Furthermore, it allows clinicians to establish accurate treatment plans (5,6). Although cone-beam computed tomography (CBCT) images might provide excellent skeletal representations, it is not possible to obtain detailed dental and interocclusal data (7,8). The quality of the image obtained through CBCT is insufficient for an accurate representation of the teeth. Therefore, the dental part in the tomographic image needs to be replaced with another modality of image to produce an accurate craniodental model (9-12).

When scanning was introduced in the world of orthognathic orthodontics/surgery, digital dental models could replace the traditional plaster ones. Such scanning of plaster models may be made through 3D scanner or CBCT (13). Then, the file to be generated from scanning is saved in DICOM format (Digital Imaging and Communications in Medicine). However, it cannot be used to perform virtual processing of digital models, so, it is required to convert it into another format named STL (Standard Tessellation Language), which is the industry standard for rapid prototyping (14).

Once obtained, digital models offer several advantages over plaster models. Digital models may be stored electronically, which significantly reduces the storage space and the risk of physical damage. In addition, these models may be easily shared via Internet, and copies of these digital models can be generated at no cost (15,16).

Digital models have been shown to be a valid tool for VSP and to perform simple diagnostic measurements, such as the size of teeth, arc width, horizontal and vertical trespasses, arc length and Bolton ratio (16). Previous studies comparing the accuracy

between linear measurements in plaster models and digital models made from scanned dental impressions or plaster models found mean differences between 0.04 and 0.62 mm (15,17,18). So, the purpose of this study was to assess the accuracy and the degree of magnification of images of plaster models obtained through 3D scanner and CBCT, comparing them with the digital caliper, which is considered the gold standard.


## Materials and methods

This retrospective study used plaster models of 40 patients of both sexes, who underwent orthognathic surgery and made orthodontic-surgical treatment in the clinic of the Dentistry Department of State University of Maringá (UEM). The inclusion criteria used in this study were the absence of positive or negative bubbles in the plaster models and presence of teeth in perfect condition without anatomical defects. The project was approved by the Ethics Committee for Research Involving Human Beings of UEM (protocol 2.336.453).


The control group used in this study was the measurement performed on 40 plaster models by Mitutoyo caliper (Copyright Mitutoyo Sul Americana Ltda). Then, the same 40 models of the control group were scanned through 3D scanner and CBCT in order to compare the degree of distortion that occurs in each technique and determine which one is the most accurate. The models were scanned on the i-CAT Next Generation (Imaging Sciences International, Hatfield, PA, USA) device, with a reconstruction volume at 0.125 mm isometric voxel, FOV (Field of View) of 8 X 8 cm, 120 kVp tube voltage and 3-8 mA tube current, in DICOM format. This equipment is installed in the Complex of Research Support Centers (COMCAP), CTS - Central of Technology in Health, Laboratory of Image in Clinical Research (LIPC) and the scanner used was the 3Shape R700 scanner (A/S, Copenhagen, Denmark). Before measurements being made on digital models, the images obtained through CBCT were converted from DICOM to STL format, as DICOM format cannot be used to perform VSP. This conversion was performed at Dolphin Imaging & Management Solutions software 11.95 3D version (Dolphin Imaging, Chatsworth, CA). It is worth mentioning that the images obtained through the 3D scanner did not require conversion, as the appliance itself generates files in STL format.

After obtaining all the images in the STL format, all digital models were tested on the Dolphin Imaging & Management Solutions software. Transverse and vertical linear anteroposterior measurements were performed in upper and lower arches: intermolar distance (IMD) - distance between the tips of the mesiobuccal cusps of permanent molars; intercanine distance (ICD) - distance between the tips of the mesiobuccal cusps of the permanent canines; segment A (SA) - mesial of the permanent right-side first molar to the mesial of the right-side central incisor); segment B (SB) - mesial of the permanent left-side first molar to the mesial of the left-side central incisor; mesiodistal (MDD) and cervico-incisal (CID) distance of the right-side central incisor (19,20). (Fig. 1A, 1B)

For statistical analysis, it was used the Pearson correlation coefficient to perform comparisons among the examiners for all variables. In addition, to compare all linear measurements carried out in upper and lower digital models, it was used the analysis of variance (ANOVA) for repeated measurements. All hypotheses tests carried out in this study considered a 5% significance level ( $P \leq 0.05$ ).



**Figure 1.** Occlusal view of the upper digital model through CBCT in Dolphin Imaging software. **A.** Demarcation of linear measurements IMD, ICD, SA, SB and MDD.



**Figure 1B.** Demarcation of linear measurement CID.

## Results


In this study, the reproducibility and accuracy of measurements performed on plaster models through digital caliper were compared with CBCT and 3D scanner measurements through Dolphin Imaging & Management Solutions software. All models of the sample ( $n=40$ ) passed through the measurements with the digital caliper and CBCT and 3D scanner scanning.

In relation to Pearson correlation coefficient (0.9902), there was an excellent correlation among the examiners. (Table 1) Stegenga et al. (21) indicated that, if the inter-examiners reliability is high, it might be assumed that intra-examiners reliability is also high, since there are fewer confounding factors with only one examiner than among several examiners.


**Table 1.** Pearson correlation coefficient between two examiners.

	<i>CC (correlation coefficient)</i>	<i>95% confidence interval</i>
Inter-examiners	0.9902	0.9749 to 0.9962

When comparing the measurements of the caliper with both methods, Bland Altman chart showed that data dispersion was within the range. Therefore, both methods are valid to assess the maxilla and the mandible. (Figs. 2 and 3)



**Figure 2.** Bland Altman chart comparing the caliper with both methods in the maxilla. Average of Caliper and Dolphin-CBCT.



**Figure 3.** Bland Altman chart comparing the caliper with both methods in the maxilla. Average of Caliper and Dolphin-Scanner.

When comparing all linear measurements (MDD, ICD, CID, IMD, SA and SB), there was no statistically significant difference for upper and lower models. (Tables 2 and 3)

**Table 2.** Mean of distances (mm) of upper and lower models.

	<b>Caliper</b>		<b>CBCT(mm)</b>		<b>Scanner (mm)</b>		<b>p-value</b>
	<b>Mean</b>	<b>SD</b>	<b>Mean</b>	<b>SD</b>	<b>Mean</b>	<b>SD</b>	
<b>IMD</b>	50.464	3.150	51.042	3.098	51.112	3.016	0.7681
<b>ICD</b>	32.759	3.903	32.887	3.910	32.935	3.928	0.9893
<b>SA</b>	36.166	3.574	35.122	3.284	35.145	3.402	0.5483
<b>SB</b>	36.080	3.507	35.017	3.320	35.065	3.345	0.5334
<b>CID</b>	8.055	1.840	7.805	1.847	7.7675	1.823	0.8650
<b>MDD</b>	8.278	0.828	8.562	0.848	8.555	0.877	0.5078

**Table 3.** Mean distances for lower orthodontic models.

	<b>Caliper</b>		<b>CBCT (mm)</b>		<b>Scanner (mm)</b>		<b>p-value</b>
	<b>Mean</b>	<b>SD</b>	<b>Mean</b>	<b>SD</b>	<b>Mean</b>	<b>SD</b>	
<b>IMD</b>	45.564	2.540	45.688	2.648	45.815	2.608	0.955
<b>ICD</b>	26.950	2.059	27.160	1.921	31.613	1.998	0.933
<b>SA</b>	31.814	1.615	31.613	1.568	31.543	1.606	0.856
<b>SB</b>	31.875	1.657	31.648	1.555	31.618	1.581	0.857
<b>CID</b>	7.126	1.274	6.953	1.245	6.990	1.263	0.901
<b>MDD</b>	5.235	0.356	5.495	0.405	5.500	0.485	0.082

In this study, by subtracting all of the means obtained from the measurements of the caliper with the means obtained from the measurements of the models scanned through CBCT and 3D scanner, it was possible to observe low results, except in segments A and B of the higher digital models. The average difference obtained through CBCT for the segments A and B was 1.044 mm and 1.063 mm, respectively. In addition, the average difference obtained through 3D scanner of the segments A and B was 1.021 mm and 1.015 mm, respectively. (Table 4)

**Table 4.** Degree of distortion of the orthodontic models scanned of the two methods with the caliper.

	<b>Degree of distortion (%)</b>
Caliper X CBCT	2.34 %
Caliper X Scanner	2.37 %

## Discussion

According to the results of this study, the reproducibility and accuracy of measurements performed on digital models produced by scanning the prints of plaster models through CBCT and 3D scanner were compared with the measurements made in the same plaster models through digital caliper, which is considered the gold standard.

Almost all of the measurements carried out in digital models had a high correlation coefficient. De Waard et al. (19) assessed the reproducibility and accuracy of linear measurements on CBCT and 3D scanner images compared with reference measurements on plaster models and concludes that the measurements found in the images obtained through CBCT had a low inter-examiner reliability, disagreeing with this study, in which there was a great inter-examiner correlation.

Quimby et al. (15), aimed at analyzing the accuracy between linear measurements in plaster models and digital models made from scanned plaster models, observed average differences between 0.04 and 0.62 mm. In this study, by subtracting all the means obtained from the measurements of the caliper with the means obtained from the measurements of the models scanned through CBCT and 3D scanner, it was possible to observe results in compliance with the literature, except in segments A and B of the maxillary models. The average difference of segments A and B for CBCT were 1.044 mm and 1.063 mm, respectively. Yet the average difference of segments A and B for the 3D scanner was 1.021 mm and 1.015 mm, respectively. Probably, this is because of the difficulty of establishing the exact point to measure such distances in the digital caliper, because manual measurements with the digital caliper depend on the positioning of the ends of the clamp on the plaster model. However, for digital measurements, the examiner should indicate two points on the computer screen that correspond to those of the distance desired. Although the examiner may manipulate the digital model on the computer screen to choose both of them precisely, the image remains in two dimensions (22).

De Waard et al. (19) concluded that scans carried out in the 3D scanner were more accurate than those made in CBCT, but the authors reported that such difference occurs due to the definitions of the voxels used for CBCT scanning. In this study, there was no statistically significant difference between both scanning methods; however, CBCT had a lesser degree of distortion than the images generated by the 3D scanner. This is probably due to the use of a small voxel and FOV (field of view) in CBCT.

Some authors (23) concluded that dental measurements implemented in digital models may be equally accurate, more reproducible and significantly faster than the measurements manually performed in traditional plaster models. This conclusion meets the results of this study, as both CBCT and scanner were consistent with the measurements performed with the digital caliper.

Regarding plaster models, digital models have some advantages in terms of cost, time and space required (24). In addition, digital models have the ability to perform 3D measurements of tooth positions. Therefore, further research is needed to evaluate the accuracy of images of plaster models obtained through 3D scanner and CBCT.

## Conclusions

According to the methodology and the results of this study, it may be concluded that only the distances of segments A and B of the upper model were not compatible in both scanning methods (3D scanner and CBCT) with the measurements of digital caliper. In addition, CBCT had a lesser degree of distortion than the 3D scanner. However, considering all of the measurements, 3D scanner and CBCT are trustworthy to perform linear measurements on digital models and are sufficiently adequate for initial diagnosis and treatment planning, and are clinically acceptable in clinical dental practices.



### Conflict of Interest

No potential conflict of interest relevant to this article was reported.

### References

1. Stokbro K, Aagaard E, Torkov P, Bell RB, Thygesen T. Virtual planning in orthognathic surgery. *Int J Oral Maxillofac Surg* 2014;43(8):957-965.
2. Xia JJ, Shevchenko L, Gateno J, Teichgraeber JF, Taylor TD, Lasky RE, et al. Outcome study of computer-aided surgical simulation in the treatment of patients with craniomaxillofacial deformities. *J Oral Maxillofac Surg* 2011; 69(7):2014-2024.
3. Baumgaertel S, Palomo JM, Palomo L, Hans MG. Reliability and accuracy of cone-beam computed tomography dental measurements. *Am J Orthod Dentofacial Orthop* 2009;136(1):19-25.
4. De Vos W, Casselman J, Swennen GR. Cone-beam computerized tomography (CBCT) imaging of the oral and maxillofacial region: a systematic review of the literature. *Int J Oral Maxillofac Surg* 2009;38(6):609-625.
5. Chen LH, Chen WH. Three-dimensional computer-assisted simulation combining facial skeleton with facial morphology for orthognathic surgery. *Int J Adult Orthodon Orthognath Surg* 1999;14(2):140-145.
6. Motohashi N, Kuroda T. A 3D computer-aided design system applied to diagnosis and treatment planning in orthodontics and orthognathic surgery. *Eur J Orthod*. 1999;21(3):263-274.
7. Xia J, Ip HH, Samman N, Wang D, Kot CS, Yeung RW, et al. Computer-assisted three-dimensional surgical planning and simulation: 3D virtual osteotomy. *Int J Oral Maxillofac Surg* 2000;29(1):11-17.
8. Terajima M, Nakasima A, Aoki Y, Goto TK, Tokumori K, Mori N, et al. A 3-dimensional method for analyzing the morphology of patients with maxillofacial deformities. *Am J Orthod Dentofacial Orthop* 2009;136(6):857-867.
9. Gateno J, Xia J, Teichgraeber JF, Rosen A. A new technique for the creation of a computerized composite skull model. *J Oral Maxillofac Surg* 2003;61(2):222-227.
10. Uechi J, Okayama M, Shibata T, Muguruma T, Hayashi K, Endo K, et al. A novel method for the 3-dimensional simulation of orthognathic surgery by using a multi-modal image-fusion technique. *Am J Orthod Dentofacial Orthop* 2006;130(6):786-798.
11. Kim BC, Lee CE, Park W, Kang SH, Zhengguo P, Yi CK, et al. Integration accuracy of digital dental models and 3-dimensional computerized tomography images by sequential point- and surface-based markerless registration. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2010;110(3):370-378.
12. Noh H, Nabha W, Cho JH, Hwang HS. Registration accuracy in the integration of laser-scanned dental images into maxillofacial cone-beam computed tomography images. *Am J Orthod Dentofacial Orthop* 2011;140(4):585-591.

13. Fleming PS, Marinho V, Johal A. Orthodontic measurements on digital study models compared with plaster models: a systematic review. *Orthod Craniofac Res* 2011;14(1):1-16.
14. Jacobs S, Grunert R, Mohr FW, Falk V. 3D-Imaging of cardiac structures using 3D heart models for planning in heart surgery: a preliminary study. *Interact Cardiovasc Thorac Surg* 2008;7(1):6-9.
15. Quimby ML, Vig KW, Rashid RG, Firestone AR. The accuracy and reliability of measurements made on computer-based digital models. *Angle Orthod* 2004;74(3):298-303.
16. Veenema AC1, Katsaros C, Boxum SC, Bronkhorst EM, Kuijpers-Jagtman AM. Index of Complexity, Outcome and Need scored on plaster and digital models. *Eur J Orthod* 2009;31(3):281-286.
17. Keating AP, Knox J, Bibb R, Zhurov AI. A comparison of plaster, digital and reconstructed study model accuracy. *J Orthod* 2008;35(3):191-201; discussion 175.
18. Bootvong K, Liu Z, McGrath C, Hägg U, Wong RW, Bendeus M, et al. Virtual model analysis as an alternative approach to plaster model analysis: reliability and validity. *Eur J Orthod* 2010;32(5):589-595.
19. de Waard O, Rangel FA, Fudalej PS, Bronkhorst EM, Kuijpers-Jagtman AM, Breuning KH. Reproducibility and accuracy of linear measurements on dental models derived from cone-beam computed tomography compared with digital dental casts. *Am J Orthod Dentofacial Orthop* 2014;146(3):328-336.
20. Lee SM, Hou Y, Cho JH, Hwang HS. Dimensional accuracy of digital dental models from cone-beam computed tomography scans of alginate impressions according to time elapsed after the impressions. *Am J Orthod Dentofacial Orthop* 2016;149(2):287-294.
21. Vandenbroucke JP, Tijssen JGP. *Klinische Epidemiologie en biostatistiek*. Codex Med., pp. 1297-1316.
22. Camardella LT, Breuning H, Vitella OV. Are there differences between comparison methods used to evaluate the accuracy and reliability of digital models? *Dental Press J Orthod* 2017;22(1):65-74.
23. Grünheid T, Patel N, De Felipe NL, Wey A, Gaillard PR, Larson BE. Accuracy, reproducibility, and time efficiency of dental measurements using different technologies. *Am J Orthod Dentofacial Orthop* 2014;145(2):157-164.
24. Hajeer MY, Millett DT, Ayoub AF, Siebert JP. Applications of 3D imaging in orthodontics: part II. *J Orthod* 2004;31(2):154-162.