

Effect of sonic *versus* ultrasonic activation on aqueous solution penetration in root canal dentin.

Dámaris Macías,¹ Víctor Bravo¹ & Diego Echeverri.¹

Affiliations: ¹School of Dentistry, Universidad Austral de Chile, Chile.

Corresponding author: Dámaris Macías. Rudloff #1640, Valdivia, Región de Los Ríos, Chile. Phone: (56-9) 50124806. E-mail: d.macias.f@gmail.com

Receipt: 10/03/2017 **Revised:** 10/25/2017
Acceptance: 12/12/2017 **Online:** 12/12/2017

Conflict of interests: The authors declare no conflicts of interest.

Ethics approval: The study protocol and informed consent were evaluated and approved by the Ethics and Research Committee of the School of Medicine of Universidad Austral de Chile.

Funding: None.

Authors' contributions: All authors carried out the entire study, and approved the final

Acknowledgements: To those who have made this research possible, especially to Dr. Adriana Covarrubias Pinto, Professor ad honorem at the Biochemistry and Microbiology Institute of the Universidad Austral de Chile, for her enormous support, willingness to help and cooperation in obtaining photographs and statistical analysis, as well as to Dr. Cristian Rosas Méndez, Dental Surgeon, for guiding us in methodology and statistics. This research is partly based on the requirements Dámaris and Víctor had to meet to obtain their degree of Dental Surgeons at the School of Dentistry at the Universidad Austral de Chile (November 2017).

Cite as: Macías D, Bravo V & Echeverri D. Effect of sonic *versus* ultrasonic activation on aqueous solution penetration in root canal dentin. J Oral Res 2018; 7(1):24-29. doi:10.17126/joralres.2017.089

Abstract: The aim of this study was to observe the penetration of an aqueous solution into the root canal dentin under sonic activation and ultrasonic activation. **Materials and Method:** This study consisted of experimental *in vitro* research. In order to achieve a closed system, the apex of 45 single-rooted teeth was sealed with wax. The step-back technique was manually performed using a K50 apical master file and 3 groups were organized according to the protocol of the final irrigant activation: Group I: non-activated Chinese ink for 30 seconds, Group II: Chinese ink sonically activated with EndoActivator for 30 seconds, and Group III: Chinese ink ultrasonically activated with Varios 350 equipment for 30 seconds. Teeth were sectioned longitudinally, and the samples obtained were observed under a stereomicroscope at 1X magnification in order to be photographed and scanned to calculate the penetration area using the Image J software. The tinted radicular area was evaluated in relation to the total area of the root dentin. The Tukey's post-hoc test and ANOVA were used for the statistical analysis ($p < 0.05$). Results: Group I and II obtained 9.13% and 9.42% penetration respectively, while in group III the highest degree of dye infiltration was achieved (13.9%), being statistically significant ($p < 0.001$). Conclusions: Ultrasonic activation produced a significantly higher penetration of the dye when compared to conventional activation and sonic activation.

Keywords: passive ultrasonic irrigation; passive sonic irrigation; irrigation penetration.

INTRODUCTION.

Irrigation is the stage during endodontic treatment during which root canals are cleaned and chemically disinfected by means of aqueous solutions which, together with the use cutting instruments, contributes to the biochemical elimination of pulp residues and microorganisms.¹

Endodontic irrigants share certain physical properties with aqueous dyes, such as Chinese ink, which has been used in endodontic treatments to compare the sealing capacity of filling materials,² the depth of penetration of the irrigant into the simulated lateral channels³ and the marginal filtration in temporary filling materials.⁴

Sodium hypochlorite (NaOCl) was the irrigant of choice because of its effectiveness in eliminating microorganisms and degrading organic material;⁵ it has a pH of 11.6 and is also a strong oxidizing agent.¹ However, it is difficult to combine it with aqueous dyes

because of these characteristics, and it is especially difficult to observe this fluid using magnification. It is in this context that the penetration of the Chinese ink is considered analogous that of irrigation solution,³ resulting in a practical and reproducible method.

Currently, the need to activate irrigation solutions in order to enhance their properties is emphasized.⁶ One way of activation is to agitate the solution using either sonic or ultrasonic devices. The former use flexible polymer inserts at 1-10kHz frequencies, while the latter use thin metal inserts and operate at 25-30 kHz frequencies.

The effectiveness of these devices has been studied in different scenarios: the removal of dentinal smear,⁶⁻⁸ the apical extrusion of sodium hypochlorite,⁹ the penetration of cement sealant into dentinal tubules,¹⁰ the removal of calcium hydroxide from the root canal,¹¹⁻¹³ and the reduction of microbial load.^{14,15} However the results obtained are not easily comparable due to the differences in the methodologies used and the studies' aims.^{16,17}

Until the present, the evaluation of the different activation systems available has been based on the observation of specific areas of the root canal with the use of different magnification levels, while the use of sonic activation or ultrasonic activation in the dentin has not been reported. The aim of this study was to observe the effect of sonic activation and ultrasonic activation on the penetration of an aqueous solution in the dentin by assessing the condition of the entire length of the root canal.

MATERIALS AND METHODS.

Design and sample

An *in vitro* experimental study was carried out. Forty-five unirradicular teeth were used, extracted due to orthodontic or periodontal reasons, free of caries and without any kind of pulpal/periapical pathologies. All procedures were performed in accordance with applicable ethical and legal provisions.

Only unirradicular teeth with closed apex, straight roots, without any previous endodontic treatment, and those that, when examined radiographically, had only

one canal, were selected. The teeth were stored in 70% denatured alcohol from the time they were obtained until their use in the study. Samples were randomly divided into three groups of 15 teeth each.

Procedure

The apex of each tooth was sealed with wax in order to achieve a closed system. Access perforations were performed in all of the 45 teeth with round dental diamond burs #4 (SS White, USA), and round carbide burs #4 (SS White, USA). For the creation of the third cervical and the root medium, Gates Glidden burs were used (Dentsply Maillefer, Switzerland), progressively increasing the D0 diameter of the file to #50, at working length (1mm from the greater apical foramen). Each file was used with balanced strong movements and between each file, 1.5ml of NaOCI 2.5% was used to irrigate. After that, 10% EDTA was used to irrigate during 1 minute Followed by rinsing with 3ml distilled water.

Irrigation was performed with a 3ml disposable syringe and a 30G irrigation side discharge needle with an inactive tip (Terumo, Japan), which was introduced passively 1mm into the working length during instrumentation. Canals were dried with paper cones #50 (Dentsply Maillefer, Switzerland).

Finally, the endodontic access of the instrumented teeth was filled with Chinese ink using an irrigation needle until completely filling the crown access.

The dye was activated as followed; Group 1: not activated (Passive), Group II: passive sonic activation, Group III: passive ultrasonic activation. For sonic activation, an EndoActivator with a blue tip was used (Dentsply Maillefer, Switzerland), at 10kHz, and for ultrasonic activation, the Varios 350 equipment (NSK Nakanishi, Japan) was used at medium power (30 kHz), together with a 25mm 15/02 stainless steel ultrasonic file with an E10 insert (NSK).

The dye remained on the teeth for a period of 30 seconds. Subsequently, it was removed by suction and the teeth dried with N° 50 paper cones.

Teeth were placed in silicone molds and transparent Acrylic Clear Duralay (Reliance® Dental Mfg Co. USA) was added to obtain dies, which were then cut longitudinally using a saw (IsoMet™ Low Speed

Saw, Buehler, USA) with a 15HC diamond disk of the following dimensions: 4''x0.12''x½'' (Buehler, USA).

A sagittal section was performed to observe the entire root canal (Figure 1).

Measures

Samples were placed under a stereomicroscope to be observed at 1x magnification (Olympus SZ61, Arquimed, Chile) and photographed. Images of the sections were scanned to calculate the penetration area of the solution as followed: the total area of root dentin (from the amelocemental junction to the apex and from the cement line to the surface of the root canal), and the tinted radicular area using the Image


J software (NIH, USA). This way it was possible to determine the percentage of penetration of the dye in the entire length of the root canal dentin and also in each of its thirds.

Statistic analysis

Data were tabulated by one of the researchers on an electronic Google Docs spreadsheet (Google Inc., USA), and were later analyzed using the SPSS Statistics Program version 20.0 (IBM, Chicago, USA).

The ANOVA *test* was applied to the quantitative variable "dye variation" and a comparison was made between the subgroups by means of the Tukey's post-hoc *test* ($p < 0.05$).

Figure 1. Sagittal section. Illustrative images of each group.



A: passive irrigation; B: irrigation with sonic activation; C: irrigation with ultrasonic activation.

RESULTS.

Six samples (two from each group) were eliminated because it was not possible to visualize the entire root canal, so finally the research was performed with groups of 13 teeth.

The passive irrigation system showed 9.13% penetration, the sonic system 9.42%, and the ultrasonic system

13.09%, its superiority being statistically significant ($p < 0.01$, Figure 2).

When comparing the penetration of the dye of each of the activation systems according to the radicular third, the ultrasonic system was observed to be significantly superior to the passive one ($p = 0.0081$) in the three zones of the root canal (Table 1).

Figure 2. Percentage of penetration according to type of activation.




Table 1. Penetration of the dye in the thirds of the root canal according to the activation method.

Type of activation	Cervical third % (S.D.)	Middle third % (S.D.)	Apical third % (S.D.)	Total penetration % (S.D.)
Passive	11.9 (5.58)	10.4 (3.94)	5.0 (2.74)	9.13 (3.05)
Sonic	11.4 (5.8)	12.8 (7.9)	4.1 (7.15)	9.42 (6.39)
Ultrasonic	14.8 (5.2)	15.2 (6.7)	9.2 (6.97)	13.09 (5.16)

S.D.: Standard Deviation.

DISCUSSION.

In this study, the penetration capacity of an aqueous solution in the entire length of the dentin of the root canal was evaluated. The ultrasonic activation showed a dye penetration significantly higher than the sonic and the passive activation. No statistically significant difference was observed between the sonic and the passive activation.

The irrigation of the root canal is one of the critical stages of the success of endodontic therapy.¹⁸ Mechanical instrumentation has been shown to leave 35% to 40% of untreated surface,¹⁹ giving microorganisms the chance to penetrate into the dentinal tubules²⁰ even beyond 500 microns. Therefore, it is essential to maximize the benefits of the irrigant by combining solutions or activation by sonic or ultrasonic agitation.²¹

Since the introduction of activation systems, their effectiveness has been questioned in various areas due to contradictory results. Kanumuru *et al.*²² Macedo *et al.*²³ and Sáinz-Pardo *et al.*²⁴ have positioned ultrasound as the most efficient method, but other authors state that sonic agitation obtains similar or even better results.^{10,17} Pérez de Arce *et al.* compared EDTA sonic activation and ultrasonic activation for the removal of dentinal smear, obtaining no significant differences between the two groups; however Nakamura *et al.* evaluated the removal of bacteria and endotoxins from the root canal, finding that ultrasound irrigation obtained better results than passive irrigation. This difference in methodology and approach of the published studies makes it difficult to compare results and assess their direct application to clinical practice.

High subjectivity is considered an important limitation of the methods used to evaluate the effectiveness of irrigants. Examples of this aspect are studies where confocal microscopy²⁵ is used, and those where a specific location within the root canal is selected.^{3,6} Results are obtained from the observation of a very small radicular area, which assumes a high probability of bias when added to the use of qualitative/quantitative scales.

The apical third was the area with the lowest penetration of all the irrigation systems, in agreement with the results reported by Spoorthy *et al.*³ and Ghorbanzadeh *et al.*²⁶ These authors state that with a suitable protocol for the removal of adhered detritus, the penetration of the irrigant in the dentinal tubules will be greater.

It should be kept in mind that the apical third has the greatest impact on the result of endodontic treatment, given that it represents the communication route between the pulp cavity and the periapical tissues. Although ultrasound showed significantly greater penetration of the dye in the dentin of the apical third compared to the other systems, it should be considered that the sample of this study included teeth with straight roots and broad canals, where the ultrasonic rigid tip reaches without any difficulty.

This is not the case of teeth with a pronounced radicular curvature, where the ultrasonic system could not approach or reach the working length.

In spite of the above, Merino *et al.*²⁷ and Castelo-Baz

*et al.*²⁸ reported better performance of ultrasound in the removal of adhered detritus in the apical third of teeth with curved canals. Additionally, the fact of having prepared the apical third to the length of a #50 file is beneficial, since according to Torabinejad *et al.*²⁹ the correct irrigation of this area is achieved by using instruments longer than a #30 file.

In the present study, the penetration of the solution was not completely homogeneous, being null in some areas. This is probably caused by detritus that was not removed by the previous irrigation protocol, by the sectioning or by the orientation of the dentinal tubules. Therefore, results should be interpreted with caution where a specific area of the canal surface is measured.

Although the method used in this study allows for the observation of the entire root canal in a single image, a component of subjectivity persists due to the inherent limitations of a two-dimensional image. It would be interesting to use this same methodology to observe the effect of the activation systems on curved root canals and to specially observe the apical third, where anatomical conditions considerably limit the irrigation and disinfection procedures during endodontic treatment.

CONCLUSION.

All the systems studied showed less penetration of the irrigant in the apical third when compared to the cervical and middle thirds, however, ultrasonic activation obtained superior results in the entire length of the root canal.

REFERENCES.

1. Balandrano Pinal F. Soluciones para irrigación en endodoncia: hipoclorito de sodio y gluconato de clorhexidina. *Rev Cient Odontol.* 2007;3(1):11-4.
2. Shahi S, Yavari HR, Rahimi S, Eskandarinezhad M, Shakouei S, Unchi M. Comparison of the sealing ability of mineral trioxide aggregate and Portland cement used as root-end filling materials. *J Oral Sci.* 2011;53(4):517-22.
3. Spoorthy E, Velmurugan N, Ballal S, Nandini S. Comparison of irrigant penetration up to working length and into simulated lateral canals using various irrigating techniques. *Int Endod J.* 2013;46(9):815-22.
4. De Castro PH, Pereira JV Jr, Sponchiado EC, Marques AA, Garcia Lda F. Evaluation of marginal leakage of different temporary restorative materials in Endodontics. *Contemp Clin Dent.* 2013;4(4):472-5.
5. Guevara D. Efecto de diferentes concentraciones de hipoclorito de sodio como irrigante endodóntico sobre propiedades físicas de la dentina. Una revisión de la literatura. Bogotá, Colombia: Universidad Nacional de Colombia. Facultad de Odontología, Especialidad en Endodoncia; 2014.
6. Kumar VR, Bahuguna N, Manan R. Comparison of efficacy of various root canal irrigation systems in removal of smear layer generated at apical third: An SEM study. *J Conserv Dent.* 2015;18(3):252-6.
7. Mancini M, Cerroni L, Iorio L, Armellini E, Conte G, Cianconi L. Smear layer removal and canal cleanliness using different irrigation systems (EndoActivator, EndoVac, and passive ultrasonic irrigation): field emission scanning electron microscopic evaluation in an in vitro study. *J Endod.* 2013;39(11):1456-60.
8. Pérez VI, Rodríguez PA, Echeverri D. Activación Sónica Versus Ultrasonica de EDTA al 10% para Remoción de Barrillo Dentinario en el Tercio Apical del Canal Radicular. *Int J Odontostomat.* 2014;8(1):153-9.
9. Helvacioğlu Kıvanç B, Deniz Arisu H, Yanar NO, Silah HM,

- İnam R, Görgül G. Apical extrusion of sodium hypochlorite activated with two laser systems and ultrasonics: a spectrophotometric analysis. *BMC Oral Health*. 2015;15:71.
10. Generali L, Cavani F, Serena V, Pettenati C, Righi E, Bertoldi C. Effect of Different Irrigation Systems on Sealer Penetration into Dentinal Tubules. *J Endod*. 2017;43(4):652–6.
 11. Capar ID, Ozcan E, Arslan H, Ertas H, Aydinbelge HA. Effect of different final irrigation methods on the removal of calcium hydroxide from an artificial standardized groove in the apical third of root canals. *J Endod*. 2014;40(3):451–4.
 12. Gokturk H, Ozkocak I, Buyukgebiz F, Demir O. Effectiveness of various irrigation protocols for the removal of calcium hydroxide from artificial standardized grooves. *J Appl Oral Sci*. 2017;25(3):290–9.
 13. Lloyd A, Navarrete G, Marchesan MA, Clement D. Removal of calcium hydroxide from Weine Type II systems using photon-induced photoacoustic streaming, passive ultrasonic, and needle irrigation: a microcomputed tomography study. *J Appl Oral Sci*. 2016;24(6):543–8.
 14. Nakamura VC, Pinheiro ET, Prado LC, Silveira AC, Carvalho APL, Mayer MPA, Gavini G. Effect of ultrasonic activation on the reduction of bacteria and endotoxins in root canals: a randomized clinical trial. *Int Endod J*. 2017:[Epub ahead of print].
 15. Neuhaus KW, Liebi M, Stauffacher S, Eick S, Lussi A. Antibacterial Efficacy of a New Sonic Irrigation Device for Root Canal Disinfection. *J Endod*. 2016;42(12):1799–1803.
 16. Mitchell RP, Baumgartner JC, Sedgley CM. Apical extrusion of sodium hypochlorite using different root canal irrigation systems. *J Endod*. 2011;37(12):1677–81.
 17. Khalap ND, Kokate S, Hegde V. Ultrasonic versus sonic activation of the final irrigant in root canals instrumented with rotary/reciprocating files: An in-vitro scanning electron microscopy analysis. *J Conserv Dent*. 2016;19(4):368–72.
 18. Gu LS, Kim JR, Ling J, Choi KK, Pashley DH, Tay FR. Review of contemporary irrigant agitation techniques and devices. *J Endod*. 2009;35(6):791–804.
 19. Peters OA, Schönenberger K, Laib A. Effects of four Ni-Ti preparation techniques on root canal geometry assessed by micro computed tomography. *Int Endod J*. 2001;34(3):221–30.
 20. Ávila S, Rosas G, García Salmones JA, Rosas Bernal N, Llamas Hernández E. Estudio histológico descriptivo de la colonización de bacterias en los túbulos dentinarios de dientes extraídos con necrosis pulpar. *Rev ADM*. 2017;74(2):69–73.
 21. Plotino G, Cortese T, Grande NM, Leonardi DP, Di Giorgio G, Testarelli L, Gambarini G. New Technologies to Improve Root Canal Disinfection. *Braz Dent J*. 2016;27(1):3–8.
 22. Kanumuru PK, Sooraparaju SG, Konda KR, Nujella SK, Reddy BK, Penigalapati SR. Comparison of Penetration of Irrigant Activated by Traditional Methods with A Novel Technique. *J Clin Diagn Res*. 2015;9(11):ZC44–7.
 23. Macedo R, Verhaagen B, Rivas DF, Versluis M, Wesselink P, van der Sluis L. Cavitation measurement during sonic and ultrasonic activated irrigation. *J Endod*. 2014;40(4):580–3.
 24. Sáinz-Pardo M, Estevez R, Pablo ÓV, Rossi-Fedele G, Cisneros R. Root canal penetration of a sodium hypochlorite mixture using sonic or ultrasonic activation. *Braz Dent J*. 2014;25(6):489–93.
 25. Vadhana S, Latha J, Velmurugan N. Evaluation of penetration depth of 2% chlorhexidine digluconate into root dentinal tubules using confocal laser scanning microscope. *Restor Dent Endod*. 2015;40(2):149–54.
 26. Ghorbanzadeh A, Aminsobhani M, Sohrabi K, Chiniforush N, Ghafari S, Shamshiri AR, Noroozi N. Penetration Depth of Sodium Hypochlorite in Dentinal Tubules after Conventional Irrigation, Passive Ultrasonic Agitation and Nd:YAG Laser Activated Irrigation. *J Lasers Med Sci*. 2016;7(2):105–11.
 27. Merino A, Estevez R, de Gregorio C, Cohenca N. The effect of different taper preparations on the ability of sonic and passive ultrasonic irrigation to reach the working length in curved canals. *Int Endod J*. 2013;46(5):427–33.
 28. Castelo-Baz P, Varela-Patiño P, Cantatore G, Domínguez-Perez A, Ruíz-Piñón M, Miguéns-Vila R, Martín-Biedma B. In vitro comparison of passive and continuous ultrasonic irrigation in curved root canals. *J Clin Exp Dent*. 2016;8(4):e437–41.
 29. Torabinejad M, Khademi AA, Babagoli J, Cho Y, Johnson WB, Bozhilov K, Kim J, Shabahang S. A new solution for the removal of the smear layer. *J Endod*. 2003;29(3):170–5.